

# DECLARATION OF CONFORMITY

DATE: DECEMBER 17, 2018

We,

**PELOTON INTERACTIVE INC.,**

**ADDRESS: 125 W 25TH STREET, 11 FL, NEW YORK, NY 10001, USA**

declare under our own responsibility that the product:

**MODEL: PLTN-HRC01**

**INTENDED USE: PELOTON HEART RATE MONITOR**

Has been tested in accordance to essential protection requirements of the red - directive 2014/53/eu, medical device directive 93/42/eec and 2007/47/ec on the approximation of the laws of the member states relating to radio spectrum matters, rf exposure/health matters and found the test results indeed meet the limitation of the relevant test standard(s) listed below:

## **1. EMC**

- EN 301 489-1: V2.2.0 (2017)
- EN 301 489-17: V3.2.0 (2017)

## **2. RADIO SPECTRUM**

- EN 300 328: V2.1.1 (2016)

## **3. SAFETY**

- EN 60601-1:2006+A11:2011+A1:2013+A12:2014

## **3. SAR**

- EN 62479 (2010)

Responsible for making this declaration is the :

**MANUFACTURER/ COMPANY NAME: PELOTON INTERACTIVE INC.**

**ADDRESS: 125 W 25TH STREET, 11 FL, NEW YORK, NY 10001, USA**



**TOM CORTESE, COO**

**PELOTON INTERACTIVE INC.**

[support@onepeloton.com](mailto:support@onepeloton.com)